

No:-MTL/PUR/43/CGDB/Press-Gauge/2019/08

Date: - 26/09/2019

Bidding type : **Single Bid**
Bid closing on : **16/10/2019 at 14.00 hrs. (IST)**
Bid Opening On : _____

ASSAM GAS COMPANY LIMITED invites quotation for supply of following item with quantity shown against it:

Sl No.	Item No/Item Code	Item Description	Unit	Qty
01	02044717	Pressure Gauge Detailed specification at Annexure-II	No	16

1. The bidder should quote their rates as per the FORMAT given below. The bidder should mention make and model no of the quoted item. In case of authorized dealer, the authorization certificate from the competent authority is to be submitted.

Item No./ Mat. Code	Material Description	Unit	Rate/Unit (INR) (A)	Freight (B)	GST @ % (C)	Amount/ Unit INR (FOR Duliajan) (A)+(B)+(C)=D	Qty (E)	Total Amount (INR) (D) X (E)

2. The TOTAL AMOUNT shall be in figure as well as in words. No overwriting shall be allowed. In case of discrepancy, the rates quoted in words shall be considered to be correct.
3. Sealed envelope containing the bid duly filled in as per above Format on bidders letter head should reach Dy. General Manager (Materials)-HOD, Assam Gas Company Limited, Duliajan-786602, Dist. Dibrugarh (Assam) on or before the bid closing date and time as mentioned above. If the bid closing date happens to be holiday/bandh call then the next working day will be considered as bid closing date. In case next working day is Saturday, the bid closing date shall be Monday.
4. The Sealed envelope containing the bid shall be marked at the top of the envelope with following details:-
 - a) AGCL Tender No.
 - b) Bid Closing date
 - c) Bidder's Name
 - d) Brief description of supply

The bid received without mention of above on the envelope shall not be accepted.

5. Bid received vide E-Mail/Fax/Telegram/Telex/Cable/ bid in open condition will not be accepted.
6. **Validity:** The bid must be valid for 180 (One Hundred Eighty) days from the date of opening of the bid.
7. **Delivery:** The ordered materials shall have to be delivered at AGCL's stores at Duliajan within 30 days from the date of receipt of the Purchase Order by Fax/Email etc. The bidder should quote their rate for FOR Duliajan.
8. **GUARANTEE/WARRANTY:** The supplier warrants that the goods supplied under the contract are new, unused and most recent of current models and that they incorporate all recent improvements in design and materials provided otherwise in the contract documents. The supplier further warrants that all goods supplied under this contract shall have no defect, arising from design, materials or workmanship or from any act of omission of the supplier.

Failure of the successful bidder to comply with the requirement shall constitute a breach of contract, cause for forfeiture of Bid security & other measures as per standing rules of the company.

All materials shall be supplied strictly in accordance with the specifications, data sheets, other conditions stated in the order. The supplier shall give the guarantee of material supplied for twelve months after commissioning of the materials or eighteen months from the date of the last shipment whichever is earlier.

9. **Liquidated Damage:** The completion of delivery of the ordered item as mentioned above within the schedule date of completion of delivery shall be deemed to be the essence of the contract. In case of delay in delivery of the materials within the stipulated period, unless such delay is attributable to owner (AGCL) or due to Force Majeure, there will be reduction in contract value @ 0.5% (Point Five Percent) per week for undelivered quantities of materials for each week of delay or part thereof subject to maximum of 5% of the total contract value, as liquidated damage. The Liquidated Damage charge shall be recovered from the supplier's bill / invoice.
10. **Payment:** 100% (Hundred Percent) payment shall be made to the supplier within 30 (Thirty) days after receipt and acceptance of materials against submission of original invoice in triplicate along with all relevant documents as per P.O. and the NIT.
11. Conditional bids are liable to rejection at the sole discretion of the Company.

12. **Applicability of Law & Jurisdiction:** The Order shall be governed and interpreted in accordance with the applicable laws of India and Dibrugarh Courts in the State of Assam shall have exclusive jurisdiction.
13. The Company reserves the right of rejecting any or all bids accepting any bid in part, without assigning any reason.

Thanking You,

Yours faithfully,

For and on behalf of **MANAGING DIRECTOR**

TECHNICAL SPECIFICATION OF PRESSURE GAUGE

- 1.** Pressure Range/ Scale Range: 0 to 5 kg/ cm²
- 2.** Display: Analog
- 3.** Liquid Filled: Not required
- 4.** Dial Size: 100 mm
- 5.** Operating Temperature: - 40°C to 60°C
- 6.** Accuracy: ± 1% of span or better
- 7.** Mount/ Connection location: Bottom
- 8.** Scale Unit: kg/ cm² & psi
- 9.** Application media: Gas
- 10.** Gauge Material: SS316
- 11.** Bourdon Tube: SS316
- 12.** Type: Bourdon Type
- 13.** Dial: White Aluminum with Black Lettering with Stop Pin
- 14.** Pointer: Black Aluminum
- 15.** Case: SS 304 with Vent Plug
- 16.** Weather Protection: Weather Resistant (NEMA 3/ IP54) Dry case
- 17.** Window Gasket: Buna-N
- 18.** Window material: Polycarbonate Acrylic